

Interpolation, Correlation and NIST Standards

Spectral irradiance measurement

$$\text{Spectrum} = \text{Assignment} \bullet \frac{\text{Test}}{\text{Calibration}} \bullet \frac{\langle \text{errors} \rangle}{\sqrt{n}} \quad n \rightarrow \infty$$

Assume all calibration have been applied so that all errors are zero mean.

$$\text{Spectrum} = \text{Assignment} \bullet \frac{\text{Test}}{\text{Calibration}}$$

I will consider only the uncertainty in the result that is due to the assignment.

Spectrum = Assignment • M

Perfect measurement and
calibration
 $u_c(M)=0$

$$\frac{\Delta \text{Spectrum}}{\text{Spectrum}} = \frac{\Delta \text{Assignment}}{\text{Assignment}}$$

Relative change / Relative
uncertainty

$\Delta \% \text{ Spectrum} = \Delta \% \text{ Assignment}$

Uncertainty propagated as
percent.

Naive analysis of uncertainty

$$Y = f(X_1, X_2, \dots, X_n) \quad \text{Measurement Equation}$$

$$u_c^2(y) = \sum_{i=1}^N \left(\frac{\partial f}{\partial x_i} \right)^2 u^2(x_i) \quad (10)$$

$$u_c^2(y) = \sum_{i=1}^N [c_i u(x_i)]^2 \quad (11a)$$

$$u_i(y) \equiv |c_i| u(x_i) \quad (11b)$$

$$c_i = \frac{u_i(y)}{u(x_i)}$$

By the book

$$Lumen = f_{lm}(S_1, S_2, \dots, S_n)$$

S_i = measured value at wavelength i
 $i = 360 \text{ nm to } 830 \text{ nm}$

$$CIEx = f_{CIEx}(S_1, S_2, \dots, S_n)$$


$$CIEy = f_{CIEy}(S_1, S_2, \dots, S_n)$$

$$u_i(Lumen) = c_i \bullet u(S_i)$$

$$u_i(Lumen) = c_i \bullet u(A_i)$$

A_i = Assigned value at wavelength i
 $u(A_i)$ is interpolated

Linear Interpolation of Uncertainties




Lumen Measurement Equation(s)

$$lm = 683 \bullet \begin{pmatrix} S_{360} \\ S_{361} \\ \dots \\ S_{830} \end{pmatrix}^T \begin{pmatrix} \bar{Y}_{360} \\ \bar{Y}_{361} \\ \dots \\ \bar{Y}_{830} \end{pmatrix}$$
$$\begin{pmatrix} S_{360} \\ S_{361} \\ \dots \\ S_{830} \end{pmatrix} = \begin{pmatrix} A_{360} \\ A_{361} \\ \dots \\ A_{830} \end{pmatrix} \bullet \begin{pmatrix} T_{360} \\ T_{361} \\ \dots \\ T_{830} \end{pmatrix} \bullet \begin{pmatrix} 1/Cal_{360} \\ 1/Cal_{361} \\ \dots \\ 1/Cal_{830} \end{pmatrix}$$


lm – lumens
 Y_λ – photopic response
 S_λ – measured spectra
 A_λ – Interpolated Assignment
 T_λ – Test measurement
 Cal_λ – Measurement of Standard

$$u_i(\text{lumen}) = c_i \cdot u(A_i)$$

Uncertainty of Lumen ($k=1$)



$$u_i(\text{color}) = c_i \cdot u(A_i)$$




Uncertainty arising from the Standard - applied at each wavelength

u_c (lumen) = 0.034% ! $u_{555\text{ nm}}$ = 0.39 %

u_c (CIEx) = 0.00007 ?

u_c (CIEy) = 0.00010 ?

Second Method apply uncertainty at interpolation



Full measurement process

Assignment $_{\lambda}$ = g(N₁, N₂, ..., N₁₂) N_i = assignment at NIST wavelengths
 λ = 360 nm to 830 nm

g is nonlinear least squares fit which I will not attempt to differentiate.

Apply uncertainty to each NIST assignment in turn and propagate through the measurement and color calculations.

Relate the differences in photometric values.

Apply uncertainty before interpolation

$$Lumen = f_{lm}(S_1, S_2, \dots, S_n)$$

$$CIEx = f_{CIEx}(S_1, S_2, \dots, S_n)$$

$$CIEy = f_{CIEy}(S_1, S_2, \dots, S_n)$$

$$u_i(Lumen) = c_i \bullet u(S_i)$$

$$u_i(Lumen) = c_i \bullet u(A_i)$$

Perfect measurement $\Delta\%S = \Delta\%A$

$$A = g(NIST)$$

$$u_i(Lumen) = c_i \bullet u(N_i)$$

N_i = Assigned value at NIST wavelength i
 $u(a_i)$ is supplied

Lumen Measurement Equation(s)

$$lm = 683 \cdot \begin{pmatrix} S_{360} \\ S_{361} \\ \dots \\ S_{830} \end{pmatrix}^T \begin{pmatrix} \bar{Y}_{360} \\ \bar{Y}_{361} \\ \dots \\ \bar{Y}_{830} \end{pmatrix}$$

$$\begin{pmatrix} S_{360} \\ S_{361} \\ \dots \\ S_{830} \end{pmatrix} = \begin{pmatrix} A_{360} \\ A_{361} \\ \dots \\ A_{830} \end{pmatrix} \cdot \begin{pmatrix} T_{360} \\ T_{361} \\ \dots \\ T_{830} \end{pmatrix} \cdot \begin{pmatrix} 1/Cal_{360} \\ 1/Cal_{361} \\ \dots \\ 1/Cal_{830} \end{pmatrix}$$

$$\begin{pmatrix} A_{360} \\ A_{361} \\ \dots \\ A_{830} \end{pmatrix} = g(\lambda)$$

$$g(\lambda) = \frac{e^{c1+c2/\lambda+c3/\lambda^2+c4/\lambda^3+c5/\lambda^4}}{\lambda^5}$$

$$\begin{pmatrix} c1 \\ c2 \\ c3 \\ c4 \\ c5 \end{pmatrix} = h \begin{pmatrix} N_{360} \\ N_{370} \\ \dots \\ N_{800} \end{pmatrix}$$

$h(\lambda)$ – Polynomial Least Square Fit
accomplished without iteration

lm – lumens
 Y_λ – photopic response
 S_λ – measured spectra
 A_λ – Interpolated Assignment
 T_λ – Test measurement
 Cal_λ – Measurement of Standard
 $g(\lambda)$ – Wien's approximation
 N_λ – NIST Assignment

Measurement Equation in EXCEL




NIST Values		
nm	E	u
360	9.023	0.62%
370	11.20	0.60%
380	13.72	0.58%
390	16.61	0.56%
400	19.78	0.55%
450	40.62	0.46%
500	68.17	0.42%
555	102.2	0.39%
600	130.0	0.37%
654.6	160.6	0.35%
700	181.7	0.34%
800	212.0	0.31%

Linear Fit	44.66	3rd Order Fit	0.9921
	-4.69		0.0100
Wien's Approximation			
2.78	31.63	31.64	0.9997
2.70	31.98	31.99	0.9998
2.63	32.32	32.32	0.9999
2.56	32.64	32.64	1.0000
2.50	32.94	32.94	1.0000
2.22	34.25	34.24	1.0002
2.00	35.30	35.28	1.0003
1.80	36.23	36.21	1.0004
1.67	36.85	36.85	1.0002
1.53	37.50	37.50	1.0000
1.43	37.96	37.96	0.9999
1.25	38.78	38.80	0.9995

x	0.00004
y	0.00007
lm	0.10%


0.334	0.334
0.334	0.334
5583	5588

Interpolated			Test Source		Tristimulus Functions				
nm	Perturbed	Baseline	nm	E	E	nm	X	Y	Z
360	9.02	9.02	360	0.000	0.000	360	0.000	0.000	0.001
361	9.22	9.22	361	0.000	0.000	361	0.000	0.000	0.001
362	9.43	9.43	362	0.000	0.000	362	0.000	0.000	0.001
363	9.64	9.64	363	0.000	0.000	363	0.000	0.000	0.001
364	9.86	9.86	364	0.000	0.000	364	0.000	0.000	0.001
365	10.08	10.08	365	0.000	0.000	365	0.000	0.000	0.001
366	10.30	10.30	366	0.000	0.000	366	0.000	0.000	0.001
367	10.52	10.52	367	0.000	0.000	367	0.000	0.000	0.001
368	10.75	10.75	368	0.000	0.000	368	0.000	0.000	0.002
369	10.98	10.98	369	0.000	0.000	369	0.000	0.000	0.002
370	11.21	11.21	370	0.000	0.000	370	0.000	0.000	0.002
371	11.45	11.45	371	0.000	0.000	371	0.000	0.000	0.002
372	11.69	11.69	372	0.000	0.000	372	0.001	0.000	0.002
373	11.93	11.93	373	0.000	0.000	373	0.001	0.000	0.003
374	12.18	12.18	374	0.000	0.000	374	0.001	0.000	0.003
375	12.43	12.43	375	0.000	0.000	375	0.001	0.000	0.003
376	12.68	12.68	376	0.000	0.000	376	0.001	0.000	0.004
377	12.94	12.94	377	0.000	0.000	377	0.001	0.000	0.005
378	13.20	13.20	378	0.000	0.000	378	0.001	0.000	0.005
379	13.46	13.46	379	0.000	0.000	379	0.001	0.000	0.006
380	13.73	13.73	380	0.000	0.000	380	0.001	0.000	0.006
381	14.00	14.00	381	0.000	0.000	381	0.002	0.000	0.007
382	14.27	14.27	382	0.000	0.000	382	0.002	0.000	0.008
383	14.55	14.55	383	0.000	0.000	383	0.002	0.000	0.009
384	14.83	14.83	384	0.000	0.000	384	0.002	0.000	0.009




$$u_i(\text{lumen}) = c_i \cdot u(N_i)$$

Uncertainty Applied before Interpolation (k=1)



$$u_i(\text{color}) = c_i \cdot u(N_i)$$

Uncertainty Applied before Interpolation (k=1)




Combined Uncertainty arising from the Standard - applied before interpolation

$$u_c \text{ (lumen)} = 0.19\% \quad \times \quad u_{555\text{nm}} = 0.39 \%$$


$$u_c \text{ (CIEx)} = 0.00030 \quad ?$$

$$u_c \text{ (CIEy)} = 0.00032 \quad ?$$

Assigned value of one wavelength changed




Assigned value of all wavelengths changed




Sources of Uncertainty	250	350	450	555	655	900	1600	2000	2300	2400	
1) HTBB Temperature Uncertainty (0.43 K at 2950 K)	0.28	0.20	0.16	0.13	0.11	0.08	0.05	0.04	0.03	0.03	0.42 K
2) HTBB spectral emissivity	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	
3) HTBB spatial uniformity	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	
4) HTBB temporal stability (0.05 K / hour)	0.03	0.02	0.02	0.02	0.01	0.01	0.01	0.00	0.00	0.00	0.04 K
5) Geometric Factors in Irradiance Transfer	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	
6) Spectroradiometer Responsivity Stability	0.30	0.30	0.15	0.15	0.15	0.15	0.15	0.15	0.15	0.50	
7) Wavelength accuracy (0.05 nm)	0.29	0.13	0.06	0.04	0.02	0.000	0.010	0.010	0.010	0.010	
8) Lamp / Spectroradiometer transfer	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	0.05	
9) Lamp Current Stability	0.03	0.02	0.02	0.02	0.02	0.01	0.01	0.01	0.01	0.01	0.08 K
10) Uncertainty of the primary working standards	0.52	0.40	0.25	0.23	0.21	0.20	0.19	0.18	0.18	0.51	
11) lamp-to-lamp transfer	0.25	0.15	0.10	0.10	0.10	0.10	0.10	0.15	0.15	0.20	
12) Long-term stability of working standards	0.66	0.47	0.36	0.30	0.25	0.18	0.10	0.08	0.07	0.07	0.99 K
13) Uncertainty of the issued	0.87	0.63	0.45	0.39	0.34	0.29	0.24	0.25	0.25	0.55	
	uncorrelated	0.50	0.37	0.21	0.21	0.21	0.21	0.23	0.23	0.55	
	correlated	0.72	0.51	0.40	0.32	0.27	0.20	0.12	0.10	0.09	1.08 K


Separation of correlated and uncorrelated




Correlation's impact on Interpolation



Uncertainty Components Separated



Uncertainty Components Separated



Uncertainty arising from the Standard - Correlations applied before interpolation

$$u_c \text{ (lumen)} = 0.34\% \quad \checkmark \quad u_{555\text{nm}} = 0.39 \%$$

$$u_c \text{ (CIEx)} = 0.00019 \quad ? \quad \text{correlation}$$

$$u_c \text{ (CIEy)} = 0.00020 \quad ? \quad \text{lowered the}$$

uncertainties

Monte Carlo Method

$$lm = f(N_{350}, N_{360}, \dots, N_{800})$$

$$u_{350}(lm) = f(N_{350} \bullet (1 + u(N_{350})), N_{360}, \dots, N_{800})$$

$$u_C(lm) = \sqrt{\sum u_i^2(lm)}$$

$$lm = f(N_{350} \bullet R_{350}, N_{360} \bullet R_{360}, \dots, N_{800} \bullet R_{800})$$

$$R_i = 1 + u(N_i) \bullet Random_{NORMAL}$$

lm is then a Probability Distribution Function

$u_C(lm)$ is the standard deviation of lm

Monte Carlo Simulation

- Correlations applied before interpolation

u_c (lumen) = 0.35%

Agrees with
discreet
analysis

u_c (CIEx) = 0.00019

u_c (CIEy) = 0.00020

Color Uncertainty Relationship

